

CERTIFICATE OF COMPLIANCE

Certificate Number 2022-07-20-E527512
Report Reference E527512-D1000-1/A0/C0-UL
Date 2022-07-20

Issued to: BAUMER A/S
Applicant Company: Runetoften 19
Arhus, 8210 Denmark

Listed Company: Same as Applicant

**This is to certify that
representative samples of**

Pressure transmitter
PFMH-##.####.#####.###0#.##0#, PFMN-
##.####.#####.#####.####; where "#" may be any digit or letter

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1, 3rd Edition, May 11, 2012, Revised July 19, 2019,
CAN/CSA-C22.2 No. 61010-1(2012-05), 3rd Edition, with
revisions through 2018-11


Additional Standards: N/A

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information.

This Certificate of Compliance does not provide authorization to apply the UL Mark. Only the UL Follow-Up Services Procedure provides authorization to apply the UL Mark.

Only those products bearing the UL Mark should be considered as being UL Certified and covered under UL's Follow-Up Services.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/about/locations/>

